



## ***Candida parapsilosis* Z011, Genomic DNA**

**Part Number: 0801537DNA-1µg**

**Part Numbers of Related Products:**

Live, Titered Organism: 0801537

**FOR RESEARCH USE ONLY  
Not for *in vitro* Diagnostic Use**

### **Product Description**

Each aliquot contains 1 µg of DNA extracted from a pure culture of *Candida parapsilosis*. The identification of this organism was confirmed by rDNA sequencing. The purity of the culture was monitored by additional culturing and Gram staining to detect any contaminating bacteria. The DNA was extracted from the cells following a protocol based on the yeast protocol provided in the Qiagen® Genomic DNA Handbook and using Qiagen® Genomic DNA Buffers with a 500/G genomic tip. This control is supplied in TE Buffer and should be frozen at -20°C or below. DNA concentration and 260/280 ratios are determined using a NanoDrop ND-1000®. The extracted DNA also tested positive on an in-house real time PCR assay.

### **Intended Use**

Purified Genomic DNA is designed for use as an amplification and/or detection control for nucleic acid testing of *C. parapsilosis*. It can also be used to determine a limit of detection (LOD), in diagnostic assay development, cross-reactivity studies or genomic sequencing. Controls should be run using the same protocols as those used to amplify extracted clinical specimens.

## Precautions

- Use Universal Precautions when handling Genomic DNA.
- The material may be re-frozen after thawing. Repetitive freezing and thawing is not recommended (aliquot material if necessary).
- To avoid cross-contamination, use separate pipette tips for all reagents.

### **DO NOT USE IN HUMANS OR AS A CLINICAL DIAGNOSTIC.**

These products are intended for research, product development or manufacturing use only. These products are NOT intended for use in the manufacture or processing of injectable products subject to licensure under section 351 of the Public Health Service Act or for any other product intended for administration to humans.



**This product was manufactured in a facility whose Quality Management System is certified as being in compliance with ISO 9001:2008 and ISO 13485:2003 standards.**

ZeptoMetrix Corporation  
872 Main Street  
Buffalo, New York 14202  
Office Phone: 716-882-0920  
Fax: 716-882-0959  
Customer Service: 800-274-5487  
www.zeptometrix.com  
E-mail: support@zeptometrix.com  
custserv@zeptometrix.com

® Registered trademarks are the property of their respective owners.

Rev.: 07/2011  
PI0801537DNA-1µg