

## Caspase-Family Colorimetric Substrate Set

**CATALOG #:** K132-7-25

**LOT #:** \_\_\_\_\_

**STORAGE CONDITIONS:** Store at -20°C.

**SHELF LIFE:** 1 year under proper storage conditions

### DESCRIPTION:

Ready-to-use colorimetric substrates for members of caspase family proteases. All substrates were provided in liquid ready-to-use form.

### KIT CONTENTS:

| Concentration | Description                       | Volume | Part Number |
|---------------|-----------------------------------|--------|-------------|
| 4 mM          | Caspase-1 Substrate, Ac-YVAD-pNA  | 125 µl | K111-25-3   |
| 4 mM          | Caspase-2 Substrate, Ac-VDVAD-pNA | 125 µl | K117-25-3   |
| 4 mM          | Caspase-3 Substrate, Ac-DEVD-pNA  | 125 µl | K106-25-3   |
| 4 mM          | Caspase-5 Substrate, Ac-WEHD-pNA  | 125 µl | K123-25-3   |
| 4 mM          | Caspase-6 Substrate, Ac-VEID-pNA  | 125 µl | K115-25-3   |
| 4 mM          | Caspase-8 Substrate, Ac-IETD-pNA  | 125 µl | K113-25-3   |
| 4 mM          | Caspase-9 Substrate, Ac-LEHD-pNA  | 125 µl | K119-25-3   |

### ASSAY PROCEDURE:

1. Induce apoptosis in cells by desired method. Concurrently incubate a control culture *without* induction.
2. Count cells and pellet 1-5 x 10<sup>6</sup> cells.
3. Resuspend cells in 50 µl of chilled Cell Lysis Buffer (Cat.# 1067-100) and incubate cells on ice for 10 minutes.
4. Centrifuge for 1 min in a microcentrifuge (10,000 x g).
5. Transfer supernatant to a fresh tube and assay protein Concentration.
6. Dilute 50-200 µg protein to 50 µl Cell Lysis Buffer for each assay.
7. Add 50 µl of 2X Reaction Buffer (Cat.# 1068-20) containing 10 mM DTT (Cat.# 1201-1) to each sample.
8. Add 5 µl of the 4 mM pNA conjugated substrates (200 µM final conc.) into each tube individually and incubate at 37°C for 1-2 hour.
9. Read samples at 400- or 405-nm in a microtiter plate reader, or spectrophotometer using a 100-µl micro quartz cuvette (Sigma), or dilute sample to 1 ml with Dilution Buffer (Cat.# 1066-100, -500) and using regular cuvette (note: Dilution of the samples proportionally decreases the reading).

Fold-increase in caspase activity can be determined by comparing these results with the level of the uninduced control.

**Note:** Background reading from cell lysates and buffers should be subtracted from the readings of both induced and the uninduced samples before calculating fold increase in caspase activity.

### RELATED PRODUCTS:

#### Apoptosis Detection Kits & Reagents

- Annexin V Kits & Bulk Reagents
- Caspase Assay Kits & Reagents
- Mitochondrial Apoptosis Kits & Reagents
- Nuclear Apoptosis Kits & Reagents
- Apoptosis Inducers and Set
- Apoptosis siRNA Vectors

#### Cell Fractionation System

- Mitochondria/Cytosol Fractionation Kit
- Nuclear/Cytosol Fractionation Kit
- Membrane Protein Extraction Kit
- Cytosol/Particulate Rapid Separation Kit
- Mammalian Cell Extraction Kit
- FractionPREP Fractionation System

#### Cell Proliferation & Senescence

- Quick Cell Proliferation Assay Kit
- Senescence Detection Kit
- High Throughput Apoptosis/Cell Viability Assay Kits
- LDH-Cytotoxicity Assay Kit
- Bioluminescence Cytotoxicity Assay Kit
- Live/Dead Cell Staining Kit

#### Cell Damage & Repair

- HDAC Fluorometric & Colorimetric Assays & Drug Discovery Kits
- HAT Colorimetric Assay Kit & Reagents
- DNA Damage Quantification Kit
- Glutathione & Nitric Oxide Fluorometric & Colorimetric Assay Kits

#### Signal Transduction

- cAMP & cGMP Assay Kits
- Akt & JNK Activity Assay Kits
- Beta-Secretase Activity Assay Kit

#### Adipocyte & Lipid Transfer

- Recombinant Adiponectin, Survivin, & Leptin
- CETP Activity Assay & Drug Discovery Kits
- PLTP Activity Assay & Drug Discovery Kits
- Total Cholesterol Quantification Kit

#### Molecular Biology & Reporter Assays

- siRNA Vectors
- Cloning Insert Quick Screening Kit
- Mitochondrial & Genomic DNA Isolation Kits
- 5 Minutes DNA Ligation Kit
- 20 Minutes Gel Staining/Destaining Kit
- β -Galactosidase Staining Kit & Luciferase Reporter Assay Kit

#### Growth Factors and Cytokines

- Monoclonal and Polyclonal Antibodies

**FOR RESEARCH USE ONLY! Not to be used on humans.**

## GENERAL TROUBLESHOOTING GUIDE FOR CASPASE COLORIMETRIC AND FLUOROMETRIC KITS:

| Problems   | Cause  | Solution  |
|--|--|---|
| Assay not working  | <ul style="list-style-type: none"> <li>• Cells did not lyse completely</li> <li>• Experiment was not performed at optimal time after apoptosis induction</li> <li>• Plate read at incorrect wavelength</li> <li>• Old DTT used</li> </ul>  | <ul style="list-style-type: none"> <li>• Resuspend the cell pellet in the lysis buffer and incubate as described in the datasheet</li> <li>• Perform a time-course induction experiment for apoptosis</li> <li>• Check the wavelength listed in the datasheet and the filter settings of the instrument</li> <li>• Always use freshly thawed DTT in the cell lysis buffer</li> </ul>  |
| High Background  | <ul style="list-style-type: none"> <li>• Increased amount of cell lysate used</li> <li>• Increased amounts of components added due to incorrect pipetting</li> <li>• Incubation of cell samples for extended periods</li> <li>• Use of expired kit or improperly stored reagents</li> <li>• Contaminated cells</li> </ul>  | <ul style="list-style-type: none"> <li>• Refer to datasheet and use the suggested cell number to prepare lysates</li> <li>• Use calibrated pipettes</li> <li>• Refer to datasheet and incubate for exact times</li> <li>• Always check the expiry date and store the individual components appropriately</li> <li>• Check for bacterial/ yeast/ mycoplasma contamination</li> </ul>   |
| Lower signal levels  | <ul style="list-style-type: none"> <li>• Cells did not initiate apoptosis</li> <li>• Very few cells used for analysis</li> <li>• Use of samples stored for a long time</li> <li>• Incorrect setting of the equipment used to read samples</li> <li>• Allowing the reagents to sit for extended times on ice</li> </ul>   | <ul style="list-style-type: none"> <li>• Determine the time-point for initiation of apoptosis after induction (time-course experiment)</li> <li>• Refer to datasheet for appropriate cell number</li> <li>• Use fresh samples or aliquot and store and use within one month for the assay</li> <li>• Refer to datasheet and use the recommended filter setting</li> <li>• Always thaw and prepare fresh reaction mix before use</li> </ul>  |
| Samples with erratic readings  | <ul style="list-style-type: none"> <li>• Uneven number of cells seeded in the wells</li> <li>• Samples prepared in a different buffer</li> <li>• Adherent cells dislodged and lost at the time of experiment</li> <li>• Cell/ tissue samples were not completely homogenized</li> <li>• Samples used after multiple freeze-thaw cycles</li> <li>• Presence of interfering substance in the sample</li> <li>• Use of old or inappropriately stored samples</li> </ul> | <ul style="list-style-type: none"> <li>• Seed only equal number of healthy cells (correct passage number)</li> <li>• Use the cell lysis buffer provided in the kit</li> <li>• Perform experiment gently and in duplicates/triplicates; apoptotic cells may become floaters</li> <li>• Use Dounce homogenizer (increase the number of strokes); observe efficiency of lysis under microscope</li> <li>• Aliquot and freeze samples, if needed to use multiple times</li> <li>• Troubleshoot as needed</li> <li>• Use fresh samples or store at correct temperatures until use</li> </ul> |
| Unanticipated results  | <ul style="list-style-type: none"> <li>• Measured at incorrect wavelength</li> <li>• Cell samples contain interfering substances</li> </ul>  | <ul style="list-style-type: none"> <li>• Check the equipment and the filter setting</li> <li>• Troubleshoot if it interferes with the kit (run proper controls)</li> </ul>  |
| General issues   | <ul style="list-style-type: none"> <li>• Improperly thawed components</li> <li>• Incorrect incubation times or temperatures</li> <li>• Incorrect volumes used</li> <li>• Air bubbles formed in the well/tube</li> <li>• Substituting reagents from older kits/ lots</li> <li>• Use of a different 96-well plate</li> </ul>   | <ul style="list-style-type: none"> <li>• Thaw all components completely and mix gently before use</li> <li>• Refer to datasheet &amp; verify the correct incubation times and temperatures</li> <li>• Use calibrated pipettes and aliquot correctly</li> <li>• Pipette gently against the wall of the well/tubes</li> <li>• Use fresh components from the same kit</li> <li>• Fluorescence: Black plates; Absorbance: Clear plates</li> </ul>   |
| <b>Note:</b> The most probable cause is listed under each section. Causes may overlap with other sections. |  |   |