

# Phosphate Colorimetric Assay Kit

(Catalog #K410-500; 500 assays; Store Kit at Room Temp.)

## I. Introduction:

Phosphate is one of the most important of the inorganic ions in biological systems. It functions in a variety of roles. One of the most important roles is as a molecular switch, turning enzyme activity on and off through the mediation of the various protein kinases and phosphatases in biological systems. Phosphate is also of great importance in mineralization processes and is a primary stimulus of algal blooms frequently found in bodies of fresh water, due to run-off from areas of high fertilizer use. The newly designed Phosphate Colorimetric Assay Kit provides an easy, quick and sensitive means of assessing phosphate over a wide range of concentrations. The assay utilizes a proprietary formulation of malachite green and ammonium molybdate which forms a chromogenic complex with phosphate ion giving an intense absorption band around 650 nm. Phosphate concentrations between 1  $\mu$ M and 1 mM, with a lower limit of detection of approximately 0.1 nmol, can be directly determined. The Phosphate Colorimetric Assay Kit provides 500 assays using microtiter plates or 100 assays using 1 ml cuvettes.

## II. Kit Contents:

Components	K410-100	Cap Code	Part Number
Phosphate Reagent	15 ml	WM	K410-100-1
Phosphate Standard (10 mM)	0.5 ml	Yellow	K410-100-2

## III. Reconstitution of Reagents:

**Phosphate Reagent:** Ready to use as supplied and may be kept at room temperature. There may be a small amount of precipitate visible which doesn't affect the assay.

## IV. Assay Protocol:

**1. Phosphate Standard Curve:** Dilute 10  $\mu$ l of the 10 mM Phosphate Standard to 990  $\mu$ l dH<sub>2</sub>O, mix well to generate 100  $\mu$ M working Phosphate Standard. Add 0, 10, 20, 30, 40, 50  $\mu$ l of the 100  $\mu$ M working Phosphate Standard to individual wells. Adjust the volume to 200  $\mu$ l with dH<sub>2</sub>O to generate 0, 1, 2, 3, 4, 5 nmol of Phosphate standard.

**2. Preparation of sample:** No sample pretreatment is necessary. Add between 0-200  $\mu$ l of sample for the assays and bring the well volume to 200  $\mu$ l with distilled water. If the approximate phosphate concentration is not known, we recommend widely different sample volumes (1, 10, 100  $\mu$ l) be tested.

The absorbance of samples should be in the linear range of the standard curve (0-5 nmol/well). If they fall outside of this range, samples should be diluted and rerun or smaller sample volumes be used. The detection limit of the assay is approximately 0.1 nmol per well (1  $\mu$ M) of Phosphate.

### 3. Reaction:

- 1) Add 30  $\mu$ l Phosphate Reagent to all standard and sample wells, mix well.
- 2) Cover the plate and incubate at room temperature for 30 min.
- 3) Read the absorbance at 650 nm using a plate reader. The color is stable for several hrs.

**4. Protocol for using 1.0 ml cuvettes:** Increase all reaction components 5X when using 1 ml cuvettes. The 1 ml total reaction mixture will contain 0-25 nmol phosphate (0-500  $\mu$ l), 150  $\mu$ l of Phosphate Reagent and made up to 1.0 ml with distilled water. Incubate at room temperature for 30 min then read at 650 nm.

## 5. Calculations:

**1) Plot standard curve:** Plot absorbance at 650 nm as a function of Phosphate concentration.

**2) Determine sample Phosphate concentration:**

$$\text{Phosphate concentration} = \frac{(\text{sample absorbance} - \text{blank absorbance})}{(\text{slope of standard curve}) \times (\mu\text{l of sample})}$$


OR

$$\text{Sa/Sv} = \text{nmol}/\mu\text{l or mM Phosphate}$$

Where Sa is the sample amount (in nmoles) read from the standard curve.

Sv is the sample volume (undiluted) added to the wells.

**Caution:** Many laboratory detergents contain high amounts of phosphates which can adhere to cleaned glassware. It is highly recommended to use disposable plastic labware for all samples, standards and reagents to avoid contamination.



**Phosphate Standard Curve:** Assay is performed following the kit protocol.

## RELATED PRODUCTS:

Apoptosis Detection Kits & Reagents  
 Glucose and Sucrose Assay Kits  
 Glutathione Assay Kits  
 NAD/NADH and NADP/NADPH Assay Kits  
 Pyruvate Assay Kits  
 Triglyceride, Fatty Acid Assay Kits  
 Inorganic ions (Na, K, Ca, Cu, Fe, Mg, Mn)

Cell Fractionation Kits  
 Cholesterol, LDL/HDL Assay Kit  
 Ethanol and Uric Acid Assay Kits  
 Lactate Assay Kits  
 Total Antioxidant Assay Kit  
 cAMP/cGMP Kits  
 Phosphatase/Kinase Assays

**FOR RESEARCH USE ONLY! Not to be used on humans.**

**GENERAL TROUBLESHOOTING GUIDE:**

<b>Problems</b>	<b>Cause</b>	<b>Solution</b>
Assay not working	<ul style="list-style-type: none"> <li>• Use of ice-cold reagent</li> <li>• Omission of a step in the protocol</li> <li>• Plate read at incorrect wavelength</li> <li>• Use of a different 96-well plate</li> </ul>	<ul style="list-style-type: none"> <li>• Reagent must be at room temperature</li> <li>• Refer and follow the data sheet precisely</li> <li>• Check the wavelength in the data sheet and the filter settings of the instrument</li> <li>• Fluorescence: Black plates ; Luminescence: White plates ; Colorimeters: Clear plates</li> </ul>
Samples with erratic readings	<ul style="list-style-type: none"> <li>• Use of an incompatible sample type</li> <li>• Samples used after multiple free-thaw cycles</li> <li>• Presence of interfering substance in the sample</li> <li>• Use of old or inappropriately stored samples</li> </ul>	<ul style="list-style-type: none"> <li>• Refer data sheet for details about incompatible samples</li> <li>• Aliquot and freeze samples if needed to use multiple times</li> <li>• Troubleshoot if needed, deproteinize samples</li> <li>• Use fresh samples or store at correct temperatures till use</li> </ul>
Lower/ Higher readings in Samples and Standards	<ul style="list-style-type: none"> <li>• Improperly thawed components</li> <li>• Use of expired kit or improperly stored reagents</li> <li>• Allowing the reagents to sit for extended times on ice</li> <li>• Incorrect incubation times or temperatures</li> <li>• Incorrect volumes used</li> </ul>	<ul style="list-style-type: none"> <li>• Thaw all components completely and mix gently before use</li> <li>• Always check the expiry date and store the components appropriately</li> <li>• Always thaw and prepare fresh reaction mix before use</li> <li>• Refer datasheet &amp; verify correct incubation times and temperatures</li> <li>• Use calibrated pipettes and aliquot correctly</li> </ul>
Readings do not follow a linear pattern for Standard curve	<ul style="list-style-type: none"> <li>• Use of partially thawed components</li> <li>• Pipetting errors in the standard</li> <li>• Pipetting errors in the reaction mix</li> <li>• Air bubbles formed in well</li> <li>• Standard stock is at an incorrect concentration</li> <li>• Calculation errors</li> <li>• Substituting reagents from older kits/ lots</li> </ul>	<ul style="list-style-type: none"> <li>• Thaw and resuspend all components before preparing the reaction mix</li> <li>• Avoid pipetting small volumes</li> <li>• Prepare a master reaction mix whenever possible</li> <li>• Pipette gently against the wall of the tubes</li> <li>• Always refer the dilutions in the data sheet</li> <li>• Recheck calculations after referring the data sheet</li> <li>• Use fresh components from the same kit</li> </ul>
Unanticipated results	<ul style="list-style-type: none"> <li>• Measured at incorrect wavelength</li> <li>• Samples contain interfering substances</li> <li>• Use of incompatible sample type</li> <li>• Sample readings above/below the linear range</li> </ul>	<ul style="list-style-type: none"> <li>• Check the equipment and the filter setting</li> <li>• Troubleshoot if it interferes with the kit</li> <li>• Refer data sheet to check if sample is compatible with the kit or optimization is needed</li> <li>• Concentrate/ Dilute sample so as to be in the linear range</li> </ul>
<b>Note:</b> The most probable list of causes is under each problem section. Causes/ Solutions may overlap with other problems.		